Insta-Test® ANALYTIC

5-WAY

Nitrate, Nitrite, pH, Alkalinity and Total Hardness

25 Test Strips

Code 3038-G

LaMotte

Immerse strip. SWIRL 2 TIMES. Remove with pads face up. DO NOT SHAKE OFF EXCESS WATER. Immediately read Hardness > Alkalinity. Wait 30 seconds, read pH > Nitrite > Nitrate.

- Keep wet fingers out of the vial
- For best results read in natural light
- Close vial tightly after removing strip
- Store in a cool, dry place

LaMotte Company
Chestertown, Maryland 21620 USA
www.lamotte.com

NO_3 Nitrate ppm

Fresh & Salt	0	OK 20	OK 40	80	160	200

NO_2 Nitrite ppm

Fresh & Salt	0	OK 0.5	1	3	5	10

pH

Fresh	6.0	6.5	OK 7.0	OK 7.5	OK 8.0	8.5	9.0
Salt	6.0	6.5	OK 7.0	OK 7.5	OK 8.0	8.5	9.0

Alk Alkalinity ppm

Fresh & Salt	0	40	80	OK 120	180	240

TH Hardness ppm

Fresh	0	OK 30	OK 60	OK 120	180	

Made in USA

6 37395 46306 8

Drop holding box and dotted tolerance line