

QUANTOFIX® Chlorine Sensitive

en

Description:

QUANTOFIX® Chlorine Sensitive are test strips for the semiquantitative determination of total chlorine in solutions. This test is also suitable for total chlorine residue control in dialysis equipment.

Measuring range:

Visually
0.1–10 mg/L Cl₂

Reflectometrically*
0.1–10 mg/L Cl₂

Color gradation:

0 · 0.1 · 0.5 · 1 · 3 · 10 mg/L Cl₂

* Together with the QUANTOFIX® Relax, these test strips may not be used for any medical application.

Pack content:

1 aluminum container with 100 test strips

Hazard warnings:

This test does not contain hazardous substances that must be labelled.

Reaction's principle:

Total chlorine reacts with the potassium iodide and the organic redox indicator contained in the test field, forming a green color.

General indications:

Remove only as many test strips as are required. Close the container immediately after removing a strip. Do not touch the test field.

Instructions for use:

1. Dip the test strip into the sample solution, moving back and forth for 15 sec.
2. Shake off excess liquid.
3. Compare the test field with the color scale. Take the value which matches closest with the colored test field (reading accuracy: ± ½ colored field of the scale).

The reaction color of the test field may change after the value has been taken. It is therefore crucial to evaluate the coloration within the prescribed time scale in order to achieve a correct result.

Quality control:

To check the correct functioning of the test strips, use a total chlorine solution with a concentration of 1 mg/L. For this purpose, first prepare a stock solution of 100 mg/L by adding 0.1 g calcium hypochlorite to 1000 mL distilled water while continuously stirring. Then, filter the solution and subject it to photometric analysis (i.e. with NANOCOLOR® Chlorine/Ozone 2, REF 985 017) to determine the exact content. Dilute 1 mL from the stock solution (100 mg/L) in 100 mL distilled water, then add 1 spatula tip of ammonium chloride (= 1 mg/L Cl₂). Immediately perform the measurement with the test strip. If the control solution produces a negative result even after repeating the process, then the remaining unused test strips must be discarded. Even during a negative control (inserting a test strip into distilled water), no positive coloration may occur. Possible reasons for incorrect functioning of the test strips may be that the use-by-date has been exceeded, the container has been left open for too long or has been stored incorrectly.

Interferences:

The presence of other strong oxidants such as bromine, iodine and hydrogen peroxide will also lead to false positive results.

Disposal:

Used test strips can be placed in the normal household waste.

Storage:




Avoid exposing the strips to sunlight and moisture. Keep container cool and dry (storage temperature not above +30 °C).

If correctly stored, the test strips may be used until the use-by-date printed on the packaging.

Additional information:

The test strip container stopper contains a non-toxic drying agent. If swallowed, drink plenty of water.

U.S. Patent No. 9,052,303



| | | | |
|--|---|--|--|
| | Diese Teststreifen entsprechen der Richtlinie für Medizinprodukte 93/42/EWG / These test strips conform to the Medical Device Directive 93/42/EEC. / Ces languettes test sont conformes à la directive 93/42/CEE relative aux dispositifs médicaux / Tiras reactivas fabricadas en conformidad con la directiva de productos sanitarios 93/42/CEE. / Deze teststrips voldoen aan de richtlijn voor medische producten 93/42/EEG. / Le presenti strisce reattive sono conformi alla direttiva 93/42/CEE sui dispositivi medici. / Te paski testowe są zgodne z dyrektywą dla wyrobów medycznych 93/42/EWG. | | |
| | Nicht wiederverwenden / Do not reuse / Ne pas réutiliser / Producto de un solo uso / Niet hergebruiken / Non riutilizzare / Nie używać ponownie | | Hersteller / Manufacturer / Fabricant / Fabricante / Fabrikant / Fabricatore / Producent |
| | Artikelnummer / Item number / Référence produit / Referencia / Artikelnummer / Numero di catalogo / Numer artykułu | | Chargencode / Batch identification / Numéro de lot / Código de lote / Productienummer / Codice del lotto / Numer partii |
| | Achtung! / Warning! / Attention! / ¡Atención! / Let op! / Attenzione! / Uwaga! | | Verwendbar bis / Use by / À utiliser avant / Fecha de caducidad / Te gebruiken tot / Utilizzare entro (anno / mese) / Przydatność do użycia |
| | Temperaturbegrenzung / Permitted storage temperature range / Limites de température / Límites de temperatura / Temperatuurbegrenzing / Limite de temperatura / Ograniczenie temperatury | | Trocken aufbewahren / Store in a dry place / Conserver au sec / Mantener seco / Op een droge plaats bewaren / Mantenere asciutto / Przechowywać w suchym miejscu |
| | Gebrauchsanweisung beachten / Please read instructions for use / Respecter les instructions d'utilisation / Obsérvense las instrucciones de uso / Lees de bijsluiting / Consultare le istruzioni per l'uso / Przestrzegać instrukcję użycia | | Packung geschlossen halten / Keep container closed / Refermer la boîte / Mantenga el envase cerrado / Verpakking gesloten houden / Conservare la confezione chiusa / Przechowywać pojemnik szczelnie zamknięty |
| | Ausreichend für <n> Prüfungen / Contains sufficient for <n> tests / Contenu suffisant pour <n> tests / Contenido suficiente para <n> tests / Voldoende voor <n> tests / Contenuto sufficiente per <n> test / Wystarczająco dla kontroli <n> | | |

